



AGILE, ROBUST,
SMART LAB SYSTEM.
SUSTAINABILITY AND RESILIENCE.

**FRICION
COEFFICIENT
TEST EQUIPMENT
& PEELING
DEVICE**

Standards

ISO8295, ISO15359, ISO638

ASTMD1894, GB/T10006, ASTM D3330

Friction Coefficient Test Equipment has been developed for the calculation of static and dynamic coefficients of friction materials compliance with ISO8295, ISO15359, ISO638, ASTMD1894, GB/T10006 standards.



+90 212 924 56 52

www.alarge.com.tr

ALARGE ALIÇ AĞACI MAK. ELEK. AR-GE SAN. TİC. LTD. ŞTİ.

Çifte Havuzlar Mah. Eski Londra Asfaltı Cd. YTÜ Teknopark D2 Blok No: 2B01 Davutpaşa - İstanbul/TÜRKİYE

Fax: +90 212 278 61 33 Mail: info@alarge.com.tr


FRICITION COEFFICIENT TEST EQUIPMENT & PEELING DEVICE


TECHNICAL SPECIFICATIONS

- ✓ Chromium-plated test surface.
- ✓ Static and dynamic friction calculation.
- ✓ Real - time data display.
- ✓ Reporting and USB recording.
- ✓ User - friendly interface.
- ✓ Fast and accurate results.
- ✓ 180 degree peeling capability.

TECHNICAL DATA


- ✓ Compatibility for composite materials such as plastic, rubber and paper.
- ✓ mN sensitivity.
- ✓ 100 mm/min speed and ± 10 mm/min precision.
- ✓ 0.5% FS precision
- ✓ .0-500 mm/min speed.
- ✓ Friction Coefficient Device
Stroke: 200 mm
- ✓ Peeling Device Stroke: 300 mm





APPLICATIONS

- ✓ Flexible packaging
- ✓ Metal
- ✓ Oppression
- ✓ Skins
- ✓ Composites
- ✓ Linoleum
- ✓ Foil
- ✓ Paper
- ✓ Plastics
- ✓ Tyre



a large


Test Sample		Second Sample
Name	Company Name	Name
Width(mm)	Operator Name	Width(mm)
Height(mm)		Height(mm)
Age(months)		Age(months)

Tare Calibration ON

Motorized Test Angle Test




FRICION COEFFICIENT TEST EQUIPMENT & PEELING DEVICE COF TEST RESULTS






Date: 22:03:2021
Time: 21:19:44

COF Test Results

Standard:	ISO 8295
Company Name:	Alarge
Operator Name:	Umut
Testing Weight(gr):	199.46
Test Mode:	Motorized Test
Sample Name:	Sample 1
Sample Width(mm):	80.0
Sample Height(mm):	200.0
Sample Age(months):	2.0
Testing Against:	The same sample
Max Static Coefficient of Friction:	0.327
Mean Static Coefficient of Friction:	0.301
Max Dynamic Coefficient of Friction:	0.322
Mean Dynamic Coefficient of Friction:	0.297





Technical Identity:

Option to test with samples of different thickness. Suitable reporting according to standards. Real time data graph on 7" screen.

Usage Identity:

Database management, system network.